



Product Service

# CERTIFICATE

No. Z1 017955 0166 Rev. 00

**Holder of Certificate:** Omicron electronics GmbH

Oberes Ried 1  
6833 Klaus  
AUSTRIA

**Certification Mark:**



**Product:** Electronic measuring equipment

**Model(s):** CMC 500

**Parameters:**  
Rated voltage: 100-240 V  
Rated frequency: 50-60 Hz  
Rated current: 15 A  
Protection class: I

**Brand Name:** OMICRON

**Tested according to:** EN 61010-1:2010/A1:2019  
EN IEC 61010-2-030:2021/A11:2021

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the Testing, Certification, Validation and Verification Regulations of TÜV SÜD Group have to be complied. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 713172918

**Valid until:** 2029-12-05

**Date,** 2024-12-10

Siemon

( Thorsten Siemon )