



America

CERTIFICATE

No. U8 15 11 17955 077

Holder of Certificate: Omicron electronics GmbH

Oberes Ried 1
6833 Klaus
AUSTRIA

Production Facility(ies): 17955

Certification Mark:



C US

Product: Electronic measuring equipment
(Measuring, testing and signal generating equipment)

Model(s): CMC 256plus, CMC310, CMC 353, CMC 356, CMS356

Parameters:

Rated voltage:	100-240V	
Rated frequency:	50/60Hz	
Rated input current:	12-10A	
Rated output:	CMC 256plus:	4x300V/6x12,5A
	CMC 310:	3x300V/3x15A or 300V/3x32A
	CMC 353:	4x300V/3x32A
	CMC 356:	4x300V/6x32A
	CMS 356:	4x300V/6x32A
Protection Class:		I

Tested according to: UL 61010-1:2012-05
UL 61010-2-030:2012-05
CAN/CSA-C22.2 No. 61010-1:2012-05
CAN/CSA-C22.2 No. 61010-2-030:2012-05

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713041050-100

Date, 2015-11-11

Page 1 of 1

