


America

CERTIFICATE

No. U8 017955 0167 Rev. 00

Holder of Certificate: Omicron electronics GmbH
Oberes Ried 1
6833 Klaus
AUSTRIA

Certification Mark:



C US

Product: Electronic measuring equipment

Brand Name: OMICRON

Model(s): CMC 500

Parameters:
Rated voltage: 100-240 V
Rated frequency: 50-60 Hz
Rated current: 15 A
Protection class: I

Tested according to:
UL 61010-1:2012/R:2023-06
UL 61010-2-030:2018/R:2023-05
CSA C22.2 No. 61010-1:2012/U3:2023-06
CSA C22.2 No. 61010-2-030:2018

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713172918

Date, 2024-12-10

(Thorsten Siemon)