



Product Service

CERTIFICATE

No. Z1 17 05 17955 099

Holder of Certificate: Omicron electronics GmbH

 Oberes Ried 1
 6833 Klaus
 AUSTRIA

**Production
Facility(ies):**

17955


Certification Mark:

Product:

Electronic measuring equipment

Model(s):

CMC430

Parameters:

Rated voltage:	100-240V
Rated frequency:	50/60Hz
Rated power:	1000W
Protection class:	I

Tested according to:

 EN 61010-1:2010
 EN 61010-2-030:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

028-713103780-000

Valid until:

2022-05-21

Date, 2017-05-23

(Erich Thurner)

Page 1 of 1

