ZERTIFIKAT ◆ CERTIFICATE ◆ 認証証書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Product Service

CERTIFICATE

No. Z1A 15 11 17955 080

Holder of Certificate: **Omicron electronics GmbH**
Oberes Ried 1
6833 Klaus
AUSTRIA

Factory(ies): 17955

Certification Mark:

Product: **Electronic measuring equipment (Measuring, testing and signal generating equipment)**

Model(s): **CMC 256plus, CMC310, CMC 353, CMC 356, CMS356**

Parameters:

Rated voltage:	100-240V	
Rated frequency:	50/60Hz	
Rated input current:	12-10A	
Rated output:	CMC 256plus:	4x300V/6x12,5A
	CMC 310:	3x300V/3x15A or 3x32A
	CMC 353:	4x300V/3x32A
	CMC 356:	4x300V/6x32A
	CMS 356:	4x300V/6x32A
Protection class:	I	

Tested according to:
EN 61010-1:2010
EN 61010-2-030:2010
AfPS GS 2014:01 PAK

The product meets the safety and health requirements of the German Product Safety Act section 20 to 22 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. See also notes overleaf.

Test report no.: 028-713041050-100

Valid until: 2020-11-18

Date, 2015-11-24

(Erich Thurner)



Page 1 of 1